



300 µL

# Biotix<sup>®</sup> uTIP<sup>®</sup> Filter and Non-Filter Pipette Tips





#### 63300045

#### 63300052

Description	uTIP° Filter Pipette T Pure Resin	ips,	uTIP° Pipette Tips, Pure Resin	
Biotix Part #	63300045	63300052	63300059	63305430
Maximum Volume	310 µL	400 μL	400 μL	400 μL
Graduation levels	10, 50, 100	10, 50, 100	10, 50, 100	10, 50, 100
Filtered	Yes	No	No	No
Low Retention	No	No	No	No
Certified Pre-Sterile	Yes	Yes	No	Yes
BioReady	Yes	Yes	Yes	Yes
Material	Clear Polypropylene	Clear Polypropylene	Clear Polypropylene	Clear Polypropylene
Packaging	96 tips/rk 10 rk/pk	96 tips/rk 10 rk/pk	96 tips/rk 10 rk/pk	1000 tips/bag 10 bags/CS
Technical Drawings	250ul   42.57mm   MAX.VOLUME   500Ul   Sevel   10 ul   sevel		250ul   59.54mm   59.54mm   MAX VOLUME   400ul   10ul   1evel   10ul   10ul	

### **Quality Testing**

	<b>-</b>	
RNase/DNase	Products are washed in distilled water and concentrated via centrifugation. Samples are added to previously established nucleic acid standards, incubated for one hour at 37°C, and tested on a 2% gel using electrophoresis. Products must show no degradation of standards to pass. Test sensitivity is 10-7 Kunitz units/µL.	
Nucleic Acid	Products are washed in distilled water and concentrated via centrifugation. Then, samples are added to protocol specified PCR reactions and thermal cycled for 50 cycles. A 2% agarose gel electrophoresis is used to examine experimental samples, positive controls, and negative controls. To pass, product samples must show no DNA amplification. Test sensitivity is 10 ng.	
Endotoxin/ Pyrogen	Products are tested for endotoxins by using the Limulus Amebocyte Lysate (LAL) gel assay according to FDA guidelines. Test sensitivity is 0.06 EU/ml.	
Trace Metal	Products are washed in distilled water. The sample is then tested using reflectometry using a single strip test for the following metals: Ca, Cu, Fe, K, Mg, Mn and Ni. Standard solutions are used as positive controls. A reader is used to detect metals to a sensitivity of 500 mg/L.	
PCR Inhibitor	Products are tested via PCR amplification and gel electrophoresis analysis. Samples must show normal amplification to be considered free of PCR inhibitors.	
Sterilization	Products are sterilized to 10-3 sterility insurance level (SAL).	
CV Test	Each lot of Biotix product is CV tested and the resulting CV is then printed on the pack label. The volume tested is equal to the maximum volume per tip size.	

## **Proprietary Technologies**

FlexFit®	Provides flexibility on the proximal end of tips, dramatically minimizing the necessary insertion to make a secure seal.
Blade®	Blade technology reduces the frequency of hanging droplets increasing both accuracy and precision.



**BIOTIX.com**