



200 μL Biotix[®] uTIP[®] Filter and Non-Filter Pipette Tips





Description	uTIP® Filter Pipette Tips, Pure Resin		uTIP° Pipette Tips, Pure Resin	
Biotix Part #	63305408	63305415	63305422	63305429
Maximum Volume	230 μL	290 μL	290 μL	290 μL
Graduation levels	10, 50, 100	10, 50, 100	10, 50, 100	10, 50, 100
Filtered	Yes	No	No	No
Low Retention	No	No	No	No
Certified Pre-Sterile	Yes	Yes	No	No
BioReady	Yes	Yes	Yes	Yes
Material	Clear Polypropylene	Clear Polypropylene	Clear Polypropylene	Clear Polypropylene
Packaging	96 tips/rk 10 rk/pk	96 tips/rk 10 rk/pk	96 tips/rk 10 rk/pk	1000 tips/bag 10 bags/CS
Technical Drawings	FBLTSE DOPPH = 11.43mm 100UL level 50.54mm MAX.VCLUME - 2220UL 10UL level		100UL level 59.54mm 59.54mm MAC-95MME 225UL 225UL 100UL level 10 = \$0.53mm	

Quality Testing

RNase/DNase	Products are washed in distilled water and concentrated via centrifugation. Samples are added to previously established nucleic acid standards, incubated for one hour at 37°C, and tested on a 2% gel using electrophoresis. Products must show no degradation of standards to pass. Test sensitivity is 10-7 Kunitz units/ μ L.
Nucleic Acid	Products are washed in distilled water and concentrated via centrifugation. Then, samples are added to protocol specified PCR reactions and thermal cycled for 50 cycles. A 2% agarose gel electrophoresis is used to examine experimental samples, positive controls, and negative controls. To pass, product samples must show no DNA amplification. Test sensitivity is 10 ng.
Endotoxin/ Pyrogen	Products are tested for endotoxins by using the Limulus Amebocyte Lysate (LAL) gel assay according to FDA guidelines. Test sensitivity is 0.06 EU/ml.
Trace Metal	Products are washed in distilled water. The sample is then tested using reflectometry using a single strip test for the following metals: Ca, Cu, Fe, K, Mg, Mn and Ni. Standard solutions are used as positive controls. A reader is used to detect metals to a sensitivity of 500 mg/L.
PCR Inhibitor	Products are tested via PCR amplification and gel electrophoresis analysis. Samples must show normal amplification to be considered free of PCR inhibitors.
Sterilization	Products are sterilized to 10-3 sterility insurance level (SAL).
CV Test	Each lot of Biotix product is CV tested and the resulting CV is then printed on the pack label. The volume tested is equal to the maximum volume per tip size.

Proprietary Technologies

FlexFit [®]	Provides flexibility on the proximal end of tips, dramatically minimizing the necessary insertion to make a secure seal.
Blade [®]	Blade technology reduces the frequency of hanging droplets increasing both accuracy and precision.



BIOTIX.com