

10 μ L


Biotix® uTIP® Filter Pipette Tip Reload



63301922

| Description | uTIP® Filter Pipette Tips, Pure Resin |
|-------------------|--|
| Biotix Part # | 63301922 |
| Maximum Volume | 17.5 μ L |
| Graduation levels | 0.5, 2, 10 |
| Filtered | Yes |
| Low Retention | No |
| Sterilized | Yes |
| BioReady | Yes |
| Material | Clear Polypropylene |
| Packaging | Reload 96 tips/rack 10 rack/pack |

Technical Drawings



Quality Testing

| | |
|-----------------------|---|
| RNase/DNase | Products are washed in distilled water and concentrated via centrifugation. Samples are added to previously established nucleic acid standards, incubated for one hour at 37°C, and tested on a 2% gel using electrophoresis. Products must show no degradation of standards to pass. Test sensitivity is 10 ⁻⁷ Kunitz units/μL. |
| Nucleic Acid | Products are washed in distilled water and concentrated via centrifugation. Then, samples are added to protocol specified PCR reactions and thermal cycled for 50 cycles. A 2% agarose gel electrophoresis is used to examine experimental samples, positive controls, and negative controls. To pass, product samples must show no DNA amplification. Test sensitivity is 10 ng. |
| Endotoxin/ Pyrogen | Products are tested for endotoxins by using the Limulus Amebocyte Lysate (LAL) gel assay according to FDA guidelines. Test sensitivity is 0.06 EU/ml. |
| Trace Metal | Products are washed in distilled water. The sample is then tested using reflectometry using a single strip test for the following metals: Ca, Cu, Fe, K, Mg, Mn and Ni. Standard solutions are used as positive controls. A reader is used to detect metals to a sensitivity of 500 mg/L. |
| PCR Inhibitor | Products are tested via PCR amplification and gel electrophoresis analysis. Samples must show normal amplification to be considered free of PCR inhibitors. |
| Sterilization | Products are sterilized to 10 ⁻⁶ sterility insurance level (SAL). |
| CV Test | Each lot of Biotix product is CV tested and the resulting CV is then printed on the pack label. The volume tested is equal to the maximum volume per tip size. |

Proprietary Technologies

| | |
|----------|--|
| FlexFit® | Provides flexibility on the proximal end of tips, dramatically minimizing the necessary insertion to make a secure seal. |
| Blade® | Blade technology reduces the frequency of hanging droplets increasing both accuracy and precision. |



BIOTIX.com